

February 16, 2021

Mr. Gary Russell
Sage Metering, Inc.
8 Harris Ct. Bldg. D-3
Monterey, CA 93940
USA

Subject: Amendment to Paramount, models 401 and 402 (AC and DC Versions)
Listing Number E112887; MET Project Number 107145
Safety Standards:

- UL 61010-1/CSA 61010-1, Second Edition: Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements
- ANSI/ISA 12.12.01, First Edition: Nonincendive Electrical Equipment for Use in Class I and II, Division 2 and Class III, Divisions 1 and 2 Hazardous (Classified) Locations
- CSA C22.2 No. 213-M1987 (R2013), First Edition, Non-incendive Electrical Equipment for Use in Class I, Division 2 Hazardous Locations
- UL1203, Fifth Edition: Explosion-Proof and Dust-Ignition-Proof Electrical Equipment for Use in Hazardous (Classified) Locations
- CSA C22.2 No. 30:20, Explosion-Proof Enclosures for use in Class I Hazardous Locations

Dear Mr. Russell:

Congratulations on successfully completing the MET Certification process for the Paramount, models 401 and 402 (AC and DC Versions). Sage Metering may continue to apply the MET Mark on the previously identified products at this time in accordance with the MET Mark Utilization Agreement or the MET Applicant Contract. The report covering the above stated products is forthcoming.

Thank you for the opportunity to perform this service for Sage Metering. We look forward to future opportunities with your company.



The Nation's First Nationally Recognized Testing Laboratory
MET Laboratories, Inc. is accredited by OSHA and the Standards Council of Canada.

NRTL

Canadian Certification has been granted under a System 3 program as defined in ISO/IEC 17067.

HEADQUARTERS: 914 WEST PATAPSCO AVENUE • BALTIMORE, MARYLAND 21230 • PHONE (410) 354-3300 • FAX (410) 354-3313

Sincerely,

MET LABORATORIES, INC.



Nehemya Cohen
Sr. Project Engineer,
Safety Laboratory



The Nation's First Nationally Recognized Testing Laboratory
MET Laboratories, Inc. is accredited by OSHA and the Standards Council of Canada.

NRTL

Canadian Certification has been granted under a System 3 program as defined in ISO/IEC 17067.

Attachment 1
Summary of Amendments/Modifications

1. Remove Hot Tap Adapter option
2. Remove conformal coating from Critical Component list
3. Remove the following Critical Drawings
 - a. 20-0041
 - b. PCB14040701
4. Update the following Critical Drawings
 - a. 25-0097 to rev D
 - b. 20-0003 to rev B
 - c. 100-0065 to rev 29



The Nation's First Nationally Recognized Testing Laboratory
MET Laboratories, Inc. is accredited by OSHA and the Standards Council of Canada.

NRTL

Canadian Certification has been granted under a System 3 program as defined in ISO/IEC 17067.